

# Erklärung Declaration

zum Einsatz von nicht-elektrischen Geräten in  
explosionsgefährdeten Bereichen.

for the use of non-electrical devices in hazardous  
environments.

**TrigasDM GmbH**  
**Erdinger Straße 2b**  
**D-85375 Neufahrn**

erklärt, in alleiniger Verantwortung, dass die  
mechanischen Turbinen-Durchflussmesser der  
Serien

declare in sole responsibility that the mechanical  
Turbine Flowmeters Series

**TF-\*\*\*\*\_\*\*\_\***  
**DM-\*\*\*\*\_\*\*\_\***  
**DML-\*\*\*\*\_\*\*\_\***

\*) und etwaige Ausführungen

\*) and variations thereof

nicht unter die Anwendung der Richtlinie  
2014/34/EU fallen.

do not fall under directive 2017/34/EU.


Durch eine Zündgefahrenanalyse wurde  
nachgewiesen, dass die Geräte keine eigene  
potenzielle Zündquelle aufweisen. Die Produkte  
können unter Beachtung der Anweisungen in den  
zugehörigen Produktinformationen in den Zonen  
1 und 2 für Gase, Nebel und Dämpfe eingesetzt  
werden.

As part of this declaration assessment, an  
ignition hazard analysis was conducted to satisfy  
the fundamental safety and health requirements  
with the following result:  
Devices can be used in zones 1 and 2 for gases,  
fog and vapors.

Grundlage für diese Bescheinigung ist eine  
Zündgefahrenanalyse an repräsentativen  
Prüfmustern der genannten Produktfamilie.

The basis for this certificate is an ignition hazard  
analysis on representative test specimens of the  
mentioned product family.

Neufahrn, 01.05.2017

  
Dipl.-Ing. Christian Stoeckle, MBA  
CEO